



Product Name : MANI® STAINLESS BURS HARD



READ THE INSTRUCTION BEFORE USE

NON STERILE

Before using the MANI® STAINLESS BURS HARD, please see the instruction manual as indicated below.

<Warnings>

1. Only skilled dentists are allowed to use.
2. Be sure to sterilize this product for each use.
3. Do not use this product except for the dental service and treatment. Use it in accordance with the intended use.

<Contraindications and prohibitions>

1. Do not use this product for a patient who indicates sensitization and allergic reaction.

<Structures and principals>

1. This product is used by connecting to dental handpiece and rotates to grind hard materials and tissues. It is a dental bur with shank and non-sterile instrument.
2. Main materials
 - 1) Working part: Stainless steel (including nickel and chromium)
 - 2) Shank : Stainless steel (including nickel and chromium)



<Purpose of use and effect-efficacy>

This instrument is defined as a rotary grinding instrument for hard tissue such as tooth and bones made of carbon steel and stainless steel. It can also be used for grinding materials of dental prostheses such as metal, plastic, porcelain and other hard materials.

<Specifications>

Under the provisions of ISO 8325 performance test, strength of the neck must exceed specified value of our standard.

<Method of manipulating or method of use>

According to its intended use, choose the most appropriate MANI® STAINLESS BURS HARD and use it for grinding hard tissues etc..

Possible Combinations with handpiece

1. Use a dental handpiece which can hold the shank precisely.
(Diameter of Contra angle shank: ϕ 2.35mm)
(Diameter of handpiece shank: ϕ 2.35mm)
2. Use the active device which could control at allowable engine speed.
3. Use a handpiece with water spray system, if needed.

<Precautions for use>

1. To prevent infection, sterilize the product by autoclave and make sure sterilization is completed before using.
2. Use with "feather touch" and keep the allowable engine speed strictly as over speed could cause fractures and injury. (※1)
3. Follow instructions of a handpiece manufacturer strictly and insert the shank all the way to avoid its half-way chucking.
4. The handpiece should be used in accordance with its instructions.
5. Choose the most appropriate type for each case and follow the general method.
6. Before using, run the instruments outside of oral cavity to make sure there is no distortion.
7. Before using, make sure the instruments outside of oral cavity that there are no deformations, scratches, cracks and tarnishes.
8. If the head of instruments is thin, long or large, there are possibilities for breaking or twisting. Because of this, be sure to avoid using unreasonable angle and excessive pressure.
9. Wear safety glasses to protect your eyes from damage. Also, wear a dust protective mask to prevent inhaling dust.
10. Wear rubber dam etc. to avoid accidental ingestion and falling.
11. Do not use this instrument for any purposes except for listed applications above.
12. Only for use by dentists.
13. This product should be treated as medical waste when disposed.
14. Dispose the instrument if damaged or contaminated.
15. After using, wash it with medical cleaning agent and brush, then wash away foreign substances like adherent body fluids and body tissues completely.
16. Set the product to a stand when cleaned by ultrasonic cleaner to avoid working part deterioration.
17. Use this product with great care to avoid puncturing fingers because of its possession of sharp-edged part.

18. This product has possibility that be corrosive if sunk into NaOCl, EDTA and etc. for a long time.
19. This product has possibility that be corrosive if sunk into salt solution. (※1)See package label for each allowable engine speed.

<Storages and duration of use>

1. Avoid storing at high temperature, humidity and direct sunlight. Keep liquids away. Store it at room temperature.
2. Do not damage or make a pinhole to packaging materials.
3. This product is subject to be improved without previous notice. Enforce first-in first-out method for stock management.
4. Do not store the instrument under a germicidal lamp to avoid deterioration.

<Maintenances and inspections>

1. Sterilize this product by autoclave under established method and term for each use.
Method of sterilization:
Put this product in a sterilization pack(or foil) and place it on a sterilization tray, or burs stand for autoclave sterilization with reference to the following terms .
Terms for sterilization:
term(1) temperature: 121degrees Celsius time: 20 minutes or more
term(2) temperature: 126degrees Celsius time: 15 minutes or more
2. Do not use high pressure steam sterilizer which heats more than 200 degrees Celsius including drying process.
3. When reusing the instrument, wash away foreign substances completely and sterilize.
4. Regarding use of medical cleaning agent, follow the instruction manual by its manufacturer strictly.
5. Dispose the products if damaged or decreased performance.

<Packagings>

Minimum packaging unit : 6pcs./ package (all sizes)
Special packaging : quantity is listed on the package label.

<Legal manufacturer >

MANI, INC.
8-3 KIYOHARA INDUSTRIAL PARK, UTSUNOMIYA, TOCHIGI,
321-3231, JAPAN
inquiry and reference: dental department
phone: +81-28-667-8592 fax: +81-28-667-8596

<EU representative >

Obelis s.a.
Bd. Général Wahis 53,
1030 Brussels, Belgium

- ※ Always keep this document near at hand. In case of loss, please contact dental department of the manufacturer for a copy.
- ※ The specifications, structures and materials are subject to be changed without previous notices for the demands of improvement.
- ※ The contents of this instruction manual are subject to be revised without previous notices.